

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2012

Name: RITTER2  
Sterilizer ID: V2313228  
02/06/2023 02:25 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.5 F 0.0  
2:00 194.0 F 2.0  
4:00 213.8 F 7.4  
6:00 233.4 F 13.4  
8:00 235.8 F 10.3  
10:00 236.3 F 8.6  
12:00 239.5 F 9.8  
14:00 250.5 F 15.7  
16:00 263.2 F 23.0  
18:00 271.8 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.7 F 29.3  
2:00 271.8 F 28.5  
3:00 272.3 F 29.3  
4:00 272.6 F 29.2  
  
Min 271.4 F 28.4  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.7 F 0.0  
5:00 227.5 F 0.0  
10:00 234.3 F 0.0  
15:00 229.5 F 0.0  
20:00 227.5 F 0.0  
25:00 226.4 F 0.0  
30:00 221.9 F 0.0  
Drying Complete

FILLING: 2:20  
HEATING: 18:06  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 00:57:26  
CYCLE COMPLETE

02/06/2023 03:22 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_