
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002217
10:57 05/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:38
STERILIZING 3:44
136.3 C 222kPa 3:44
136.1 C 221kPa 4:14
136.4 C 226kPa 4:44
136.4 C 224kPa 5:14
136.4 C 226kPa 5:44
136.4 C 224kPa 6:14
136.4 C 223kPa 6:44
136.4 C 226kPa 7:14
136.4 C 226kPa 7:14
VENTING 7:14
AIR DRYING 8:07
CYCLE COMPLETE 8:37

Digital Signature #
ADA1100BD4C23AE0
