
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001072
16:46 11/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:20
PRESSURIZING 4:22
STERILIZING 5:37
136.0 C 220kPa 5:37
136.4 C 225kPa 6:07
136.4 C 224kPa 6:37
136.2 C 223kPa 7:07
136.2 C 223kPa 7:37
136.4 C 223kPa 8:07
136.3 C 224kPa 8:37
136.4 C 224kPa 9:07
136.4 C 224kPa 9:07
VENTING 9:07
AIR DRYING 10:05
CYCLE COMPLETE 10:50

Digital Signature #
9EE105C462BDC155
