

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2388

Name: RITTER2  
Sterilizer ID: V2313228  
05/18/2023 10:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.0 F 0.0  
2:00 180.2 F 3.0  
4:00 186.4 F 7.1  
6:00 206.7 F 11.8  
8:00 222.1 F 6.8  
10:00 229.8 F 12.1  
12:00 235.0 F 9.5  
14:00 236.2 F 9.4  
16:00 248.3 F 14.5  
18:00 259.2 F 20.8  
20:00 269.8 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.7  
1:00 272.6 F 29.9  
2:00 272.2 F 30.3  
3:00 272.1 F 29.7  
4:00 272.3 F 30.3  
  
Min 271.6 F 29.6  
Max 272.6 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.1 F 0.0  
5:00 228.3 F 0.0  
10:00 238.9 F 0.0  
15:00 230.0 F 0.0  
20:00 229.1 F 0.0  
25:00 229.0 F 0.0  
30:00 225.0 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 21:51  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:00:59  
CYCLE COMPLETE

05/18/2023 11:13 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_