
Midmark M11 - v1.0.0
Total Cycles: 664

Name: MIDMARK
Sterilizer ID: V2178198
02/11/2021 09:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.9 F 0.0
2:00 72.3 F 0.0
4:00 88.7 F 0.0
6:00 108.4 F 0.1
8:00 138.0 F 0.2
10:00 176.3 F 0.9
12:00 221.5 F 4.3
14:00 230.1 F 7.9
16:00 238.6 F 11.6
18:00 232.2 F 7.0
20:00 241.0 F 12.3
22:00 239.0 F 9.6
24:00 237.2 F 8.5
26:00 249.5 F 14.6
28:00 261.3 F 21.3
30:00 272.1 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.1 F 28.8
2:00 272.7 F 28.8
3:00 272.3 F 28.7
4:00 272.2 F 28.5

Min 271.3 F 28.3
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 233.3 F 0.0
10:00 228.1 F 0.0
15:00 213.9 F 0.0
20:00 220.8 F 0.0
25:00 236.0 F 0.0
30:00 233.0 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 31:05
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:09:02
CYCLE COMPLETE

02/11/2021 10:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____