
Midmark M11 - v1.0.5
Total Cycles: 2517

Name: RITTER2
Sterilizer ID: V2313228
06/14/2023 01:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.1 F 0.0
2:00 167.6 F 3.4
4:00 183.5 F 7.9
6:00 210.7 F 13.3
8:00 222.6 F 8.4
10:00 235.6 F 11.4
12:00 239.0 F 11.8
14:00 239.2 F 10.7
16:00 252.8 F 17.6
18:00 264.8 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.5 F 29.8
2:00 271.7 F 29.7
3:00 272.4 F 29.7
4:00 271.6 F 29.7

Min 271.5 F 29.3
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 223.5 F 0.0
10:00 234.4 F 0.0
15:00 223.7 F 0.0
20:00 218.2 F 0.0
25:00 216.4 F 0.0
30:00 214.7 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 19:48
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

06/14/2023 02:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____