

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006557  
9:57 06/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:34  
PRESSURIZING 3:35  
STERILIZING 4:23  
136.0 C 220kPa 4:23  
136.4 C 222kPa 4:53  
136.4 C 223kPa 5:23  
136.4 C 223kPa 5:53  
136.2 C 222kPa 6:23  
136.4 C 223kPa 6:53  
136.1 C 224kPa 7:23  
136.4 C 224kPa 7:53  
136.4 C 224kPa 7:53  
VENTING 7:53  
AIR DRYING 8:44  
CYCLE COMPLETE 8:48

-----  
Digital Signature #  
5DCFC8EDC9E6FC36  
-----