

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3880

Name: RITTER2  
Sterilizer ID: V2313228  
08/21/2024 02:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.0 F 0.0  
2:00 153.9 F 2.7  
4:00 165.3 F 6.0  
6:00 189.1 F 10.4  
8:00 211.8 F 5.6  
10:00 220.7 F 9.0  
12:00 229.2 F 11.7  
14:00 235.3 F 11.1  
16:00 235.8 F 9.5  
18:00 249.0 F 15.9  
20:00 260.5 F 22.2  
22:00 270.9 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.5  
1:00 271.1 F 29.7  
2:00 272.5 F 30.8  
3:00 272.6 F 30.4  
4:00 271.6 F 30.0

Min 270.9 F 29.4  
Max 273.7 F 30.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.1 F 0.0  
5:00 231.3 F 0.0  
10:00 229.2 F 0.0  
15:00 220.1 F 0.0  
20:00 220.3 F 0.0  
25:00 216.3 F 0.0  
30:00 211.1 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 23:02  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:01:47  
CYCLE COMPLETE

08/21/2024 03:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_