
Midmark M11 - v1.0.0
Total Cycles: 668

Name: MIDMARK
Sterilizer ID: V2178198
02/17/2021 07:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.2 F 0.0
2:00 84.5 F 0.0
4:00 84.4 F 0.1
6:00 94.8 F 0.1
8:00 123.8 F 0.2
10:00 211.5 F 0.8
12:00 215.3 F 4.6
14:00 231.4 F 7.6
16:00 239.6 F 11.1
18:00 232.6 F 6.9
20:00 241.9 F 11.7
22:00 237.5 F 8.9
24:00 239.6 F 9.6
26:00 248.5 F 14.0
28:00 260.9 F 20.9
30:00 271.8 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.3
1:00 272.0 F 28.5
2:00 272.3 F 28.7
3:00 272.4 F 28.8
4:00 272.5 F 28.7

Min 271.9 F 28.2
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.9 F 0.0
5:00 234.0 F 0.0
10:00 230.8 F 0.0
15:00 229.9 F 0.0
20:00 225.5 F 0.0
25:00 227.6 F 0.0
30:00 200.9 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:09
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:08:54
CYCLE COMPLETE

02/17/2021 08:43 AM
ACCEPTED: X
REJECTED:
Operator: U01