

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3532

Name: RITTER2  
Sterilizer ID: V2313228  
04/16/2024 10:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.5 F 0.0  
2:00 157.0 F 1.6  
4:00 177.5 F 5.1  
6:00 201.9 F 10.2  
8:00 215.2 F 5.1  
10:00 231.0 F 9.3  
12:00 235.4 F 9.7  
14:00 242.5 F 12.4  
16:00 241.9 F 11.4  
18:00 255.3 F 18.3  
20:00 266.9 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.5 F 29.1  
2:00 272.0 F 29.2  
3:00 272.2 F 28.9  
4:00 272.3 F 29.4

Min 271.6 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.9 F 0.0  
5:00 233.7 F 0.0  
10:00 210.1 F 0.0  
15:00 209.0 F 0.0  
20:00 207.1 F 0.0  
25:00 214.2 F 0.0  
30:00 214.0 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 21:25  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:00:10  
CYCLE COMPLETE

04/16/2024 11:34 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_