
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000243
14:33 20/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:07
STERILIZING 4:27
136.5 C 224kPa 4:27
136.2 C 223kPa 4:57
136.1 C 223kPa 5:27
136.2 C 222kPa 5:57
136.2 C 224kPa 6:27
136.3 C 224kPa 6:57
136.3 C 224kPa 7:27
136.4 C 224kPa 7:57
136.4 C 224kPa 7:57
VENTING 7:57
AIR DRYING 8:53
CYCLE COMPLETE 8:57

Digital Signature #
F0A65D092E34A01D
