
Midmark M11 - v1.0.0
Total Cycles: 1857

Name: RITTER 1
Sterilizer ID: V2178198
04/22/2022 10:52 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.5 F 0.0
2:00 154.1 F 1.7
4:00 179.2 F 5.8
6:00 202.9 F 11.3
8:00 219.7 F 5.3
10:00 235.4 F 11.4
12:00 235.5 F 8.6
14:00 238.2 F 8.9
16:00 249.1 F 14.4
18:00 262.8 F 22.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.8
1:00 271.6 F 27.7
2:00 271.8 F 27.7
3:00 272.1 F 28.0

Min 271.5 F 27.6
Max 272.9 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
Dry Cycle Aborted

FILLING: 1:51
HEATING: 19:55
STERILIZING: 3:00
VENTING: 2:37
DRYING: 2:04
TOTAL CYCLE: 00:29:27

04/22/2022 11:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____