
Midmark M11 - v1.0.5
Total Cycles: 3003

Name: RITTER2
Sterilizer ID: V2313228
10/12/2023 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.9 F 0.0
2:00 150.3 F 2.4
4:00 170.8 F 6.0
6:00 208.0 F 11.0
8:00 224.9 F 5.2
10:00 231.9 F 10.5
12:00 231.5 F 7.3
14:00 243.9 F 13.4
16:00 243.4 F 12.2
18:00 256.2 F 19.0
20:00 267.8 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.7 F 29.4
2:00 272.8 F 29.9
3:00 271.4 F 29.1
4:00 273.2 F 29.7

Min 271.3 F 29.1
Max 273.2 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.4 F 0.0
5:00 236.2 F 0.0
10:00 202.3 F 0.0
15:00 200.3 F 0.0
20:00 198.8 F 0.0
25:00 199.7 F 0.0
30:00 200.4 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:60
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

10/12/2023 02:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____