
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002596
11:09 24/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:29
STERILIZING 3:38
136.0 C 220kPa 3:38
136.4 C 225kPa 4:08
136.2 C 224kPa 4:38
136.1 C 222kPa 5:08
136.4 C 226kPa 5:38
136.1 C 224kPa 6:08
136.4 C 225kPa 6:38
136.4 C 223kPa 7:08
136.4 C 223kPa 7:08
VENTING 7:08
AIR DRYING 8:04
CYCLE COMPLETE 12:47

Digital Signature #
FB0B160A3C7995B1
