

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1925

Name: RITTER2  
Sterilizer ID: V2313228  
01/12/2023 08:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.7 F 0.0  
2:00 79.3 F 0.0  
4:00 95.6 F 0.0  
6:00 118.4 F 0.0  
8:00 161.3 F 0.1  
10:00 191.3 F 0.3  
12:00 206.1 F 3.9  
14:00 217.2 F 7.4  
16:00 227.5 F 11.2  
18:00 229.6 F 6.9  
20:00 241.9 F 12.5  
22:00 240.2 F 10.9  
24:00 238.1 F 9.7  
26:00 251.9 F 16.5  
28:00 263.4 F 23.2  
30:00 271.8 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.4 F 29.1  
2:00 271.9 F 29.2  
3:00 272.1 F 28.9  
4:00 272.1 F 29.3

Min 271.5 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.9 F 0.0  
5:00 216.8 F 0.0  
10:00 236.6 F 0.0  
15:00 229.8 F 0.0  
20:00 229.4 F 0.0  
25:00 231.7 F 0.0  
30:00 229.0 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:06  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:08:00  
CYCLE COMPLETE

01/12/2023 09:28 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_