
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001844
8:55 15/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:08
STERILIZING 3:16
136.0 C 221kPa 3:16
136.1 C 223kPa 3:46
136.2 C 223kPa 4:16
136.4 C 224kPa 4:46
136.2 C 223kPa 5:16
136.2 C 223kPa 5:46
136.3 C 225kPa 6:16
136.2 C 223kPa 6:46
136.2 C 223kPa 6:46
VENTING 6:46
AIR DRYING 7:41
CYCLE COMPLETE 8:40

Digital Signature #
3236254D1D2A43BF
