
Midmark M11 - v1.0.5
Total Cycles: 886

Name: RITTER2
Sterilizer ID: V2313228
03/16/2022 02:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.1 F 0.0
2:00 193.4 F 3.6
4:00 212.7 F 9.2
6:00 228.6 F 8.8
8:00 237.0 F 12.4
10:00 239.2 F 10.8
12:00 238.4 F 9.9
14:00 252.7 F 17.0
16:00 264.3 F 24.2
18:00 270.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 29.9
1:00 272.5 F 30.0
2:00 272.8 F 30.4
3:00 271.9 F 30.1
4:00 272.6 F 30.1

Min 271.9 F 29.8
Max 273.0 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.4 F 0.0
5:00 216.3 F 0.0
10:00 234.1 F 0.0
15:00 227.2 F 0.0
20:00 225.3 F 0.0
25:00 223.4 F 0.0
30:00 221.5 F 0.0
Drying Complete

FILLING: 2:50
HEATING: 18:47
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

03/16/2022 03:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____