
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002127
14:19 04/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:47
STERILIZING 3:59
136.4 C 224kPa 3:59
136.3 C 224kPa 4:29
136.2 C 223kPa 4:59
136.2 C 223kPa 5:29
136.4 C 223kPa 5:59
136.2 C 223kPa 6:29
136.4 C 225kPa 6:59
136.4 C 225kPa 7:29
136.4 C 225kPa 7:29
VENTING 7:29
AIR DRYING 8:20
CYCLE COMPLETE 27:51

Digital Signature #
37F08827195D409F
