
Midmark M11 - v1.0.0
Total Cycles: 4345

Name: RITTER 1
Sterilizer ID: V2178198
02/29/2024 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.0 F 0.0
2:00 163.4 F 2.3
4:00 166.4 F 5.5
6:00 188.1 F 9.8
8:00 215.2 F 5.5
10:00 227.2 F 7.8
12:00 238.4 F 12.6
14:00 235.4 F 9.1
16:00 239.0 F 9.8
18:00 246.6 F 13.4
20:00 258.7 F 19.9
22:00 269.5 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.1 F 28.3
2:00 272.3 F 28.8
3:00 271.6 F 28.5
4:00 272.0 F 28.3

Min 271.4 F 28.1
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.9 F 0.0
5:00 229.7 F 0.0
10:00 224.3 F 0.0
15:00 214.5 F 0.0
20:00 233.6 F 0.0
25:00 228.9 F 0.0
30:00 232.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 22:56
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:01:15
CYCLE COMPLETE

02/29/2024 11:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____