

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000130  
15:53 11/03/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:28  
PRESSURIZING 3:30  
STERILIZING 4:40  
136.0 C 220kPa 4:40  
136.4 C 222kPa 5:10  
136.2 C 222kPa 5:40  
136.0 C 222kPa 6:10  
136.2 C 223kPa 6:40  
136.3 C 223kPa 7:10  
136.2 C 222kPa 7:40  
136.3 C 224kPa 8:10  
136.3 C 224kPa 8:10  
VENTING 8:10  
AIR DRYING 9:04  
CYCLE COMPLETE 9:14  
-----

Digital Signature #  
E8A1004FC9DBA166  
-----