

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000087  
8:50 28/08/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:56  
PRESSURIZING 3:57  
STERILIZING 5:08  
136.0 C 221kPa 5:08  
136.1 C 221kPa 5:38  
136.4 C 225kPa 6:08  
136.4 C 224kPa 6:38  
136.4 C 222kPa 7:08  
136.1 C 221kPa 7:38  
136.4 C 221kPa 8:08  
136.3 C 224kPa 8:38  
136.3 C 224kPa 8:38  
VENTING 8:38  
AIR DRYING 9:31  
CYCLE COMPLETE 24:31  
-----

Digital Signature #  
691192E6D28A3E31  
-----