

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1081

Name: RITTER2  
Sterilizer ID: V2313228  
05/09/2022 08:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.3 F 0.0  
2:00 171.4 F 1.5  
4:00 172.2 F 5.4  
6:00 193.5 F 10.8  
8:00 223.3 F 5.6  
10:00 236.5 F 11.0  
12:00 233.5 F 8.2  
14:00 242.9 F 11.4  
16:00 245.4 F 13.0  
18:00 257.5 F 19.6  
20:00 268.2 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 271.9 F 29.2  
2:00 271.9 F 29.3  
3:00 272.0 F 29.1  
4:00 272.0 F 29.1

Min 271.6 F 29.0  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.5 F 0.0  
5:00 233.4 F 0.0  
10:00 223.9 F 0.0  
15:00 218.8 F 0.0  
20:00 192.2 F 0.0  
25:00 206.5 F 0.0  
30:00 203.9 F 0.0  
Drying Complete

FILLING: 1:41  
HEATING: 21:55  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:46  
CYCLE COMPLETE

05/09/2022 09:22 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02