
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007243
6:29 23/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:56
PRESSURIZING 3:57
STERILIZING 5:07
136.4 C 222kPa 5:07
136.4 C 224kPa 5:37
136.4 C 223kPa 6:07
136.4 C 221kPa 6:37
136.4 C 221kPa 7:07
136.4 C 222kPa 7:37
136.3 C 223kPa 8:07
136.3 C 220kPa 8:37
136.3 C 220kPa 8:37
VENTING 8:37
AIR DRYING 9:30
CYCLE COMPLETE 39:30

Digital Signature #
32F86F8B16CF1BC6
