
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000218
12:23 22/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:24
PRESSURIZING 4:24
STERILIZING 5:53
136.4 C 232kPa 5:53
136.4 C 233kPa 6:23
136.3 C 232kPa 6:53
136.4 C 232kPa 7:23
136.4 C 232kPa 7:53
136.4 C 233kPa 8:23
136.4 C 233kPa 8:53
136.4 C 233kPa 9:23
136.4 C 233kPa 9:23
VENTING 9:23
AIR DRYING 10:25
CYCLE COMPLETE 15:08

Digital Signature #
B020C67B3FE84E7F
