

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1191

Name: RITTER 1  
Sterilizer ID: V2178198  
09/22/2021 06:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.2 F 0.0  
2:00 72.9 F 0.0  
4:00 77.4 F 0.0  
6:00 101.5 F 0.0  
8:00 133.0 F 0.1  
10:00 164.7 F 0.1  
12:00 191.5 F 0.2  
14:00 212.9 F 2.7  
16:00 220.4 F 5.4  
18:00 226.1 F 8.1  
20:00 233.6 F 11.4  
22:00 232.6 F 7.0  
24:00 243.5 F 12.0  
26:00 239.1 F 9.7  
28:00 239.2 F 9.5  
30:00 248.5 F 14.1  
32:00 259.2 F 19.9  
34:00 268.9 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 27.8  
1:00 272.6 F 28.6  
2:00 272.0 F 28.0  
3:00 271.6 F 28.0  
4:00 272.6 F 28.6  
  
Min 271.4 F 27.7  
Max 272.7 F 28.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.2 F 0.0  
5:00 215.3 F 0.0  
10:00 223.9 F 0.0  
15:00 222.7 F 0.0  
20:00 213.1 F 0.0  
25:00 210.1 F 0.0  
30:00 209.6 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 35:15  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:13:31  
CYCLE COMPLETE

09/22/2021 07:54 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01