
Midmark M11 - v1.0.0
Total Cycles: 4576

Name: RITTER 1
Sterilizer ID: V2178198
05/23/2024 12:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 88.0 F 0.0
2:00 90.0 F 0.6
4:00 108.9 F 2.1
6:00 136.5 F 4.5
8:00 162.1 F 7.3
10:00 183.9 F 10.9
12:00 203.7 F 5.1
14:00 210.4 F 4.5
16:00 221.7 F 7.8
18:00 230.4 F 11.6
20:00 229.4 F 7.0
22:00 240.8 F 11.9
24:00 238.1 F 9.6
26:00 249.5 F 15.1
28:00 259.6 F 20.6
30:00 269.0 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 29.1
2:00 272.3 F 28.8
3:00 271.5 F 28.3
4:00 271.7 F 28.7

Min 271.3 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 196.7 F 0.0
10:00 195.9 F 0.0
15:00 204.0 F 0.0
20:00 208.5 F 0.0
25:00 219.1 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 31:16
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:09:41
CYCLE COMPLETE

05/23/2024 01:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____