
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
2.1uS / 1.3ppm
CYCLE NUMBER 000069
11:06 17/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:40
STERILIZING 3:58
136.4 C 223kPa 3:58
136.4 C 225kPa 4:28
136.4 C 224kPa 4:58
136.4 C 224kPa 5:28
136.4 C 223kPa 5:58
136.3 C 224kPa 6:28
136.3 C 225kPa 6:58
136.3 C 225kPa 7:28
136.3 C 225kPa 7:28
VENTING 7:28
AIR DRYING 8:20
CYCLE COMPLETE 22:37

Digital Signature #
DAB21F763860C715
