

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1349

Name: RITTER 1  
Sterilizer ID: V2178198  
11/09/2021 04:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.3 F 0.0  
2:00 196.2 F 4.3  
4:00 213.3 F 8.8  
6:00 223.5 F 8.9  
8:00 233.3 F 9.6  
10:00 235.3 F 8.8  
12:00 244.2 F 12.8  
14:00 242.4 F 11.5  
16:00 255.8 F 18.4  
18:00 267.6 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.5 F 29.1  
2:00 272.3 F 28.6  
3:00 271.1 F 28.1  
4:00 271.7 F 28.7  
  
Min 271.0 F 28.0  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.6 F 0.0  
5:00 221.4 F 0.0  
10:00 212.3 F 0.0  
15:00 213.8 F 0.0  
20:00 220.0 F 0.0  
25:00 220.7 F 0.0  
30:00 221.4 F 0.0  
Drying Complete

FILLING: 2:15  
HEATING: 19:23  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:50  
CYCLE COMPLETE

11/09/2021 05:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_