
Midmark M11 - v1.0.5
Total Cycles: 4447

Name: RITTER2
Sterilizer ID: V2313228
03/24/2025 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.2 F 0.0
2:00 161.1 F 2.0
4:00 172.7 F 5.8
6:00 195.5 F 10.1
8:00 216.4 F 7.7
10:00 224.9 F 9.8
12:00 233.7 F 9.2
14:00 240.7 F 12.5
16:00 240.1 F 11.2
18:00 252.0 F 17.5
20:00 263.0 F 24.0
22:00 272.3 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 30.5
1:00 271.9 F 30.9
2:00 272.1 F 30.9
3:00 272.0 F 30.8
4:00 271.8 F 30.8

Min 271.7 F 30.4
Max 272.5 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 231.3 F 0.0
10:00 223.0 F 0.0
15:00 217.9 F 0.0
20:00 216.6 F 0.0
25:00 215.3 F 0.0
30:00 211.6 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 22:49
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:01:58
CYCLE COMPLETE

03/24/2025 01:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____