
Midmark M11 - v1.0.0
Total Cycles: 714

Name: MIDMARK
Sterilizer ID: V2178198
03/01/2021 01:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.7 F 0.0
2:00 128.0 F 1.3
4:00 149.5 F 4.4
6:00 179.0 F 8.6
8:00 212.2 F 10.4
10:00 216.7 F 7.8
12:00 234.7 F 13.0
14:00 238.2 F 10.1
16:00 234.5 F 8.2
18:00 249.0 F 14.9
20:00 260.9 F 21.7
22:00 271.8 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 273.0 F 29.5
2:00 271.7 F 28.6
3:00 272.4 F 29.5
4:00 272.7 F 29.1

Min 271.2 F 28.5
Max 273.2 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.8 F 0.0
5:00 237.9 F 0.0
10:00 227.7 F 0.0
15:00 184.8 F 0.0
20:00 191.5 F 0.0
25:00 174.1 F 0.0
30:00 183.3 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 22:36
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:41
CYCLE COMPLETE

03/01/2021 02:17 PM
ACCEPTED: X
REJECTED:
Operator: U01