
Midmark M11 - v1.0.0
Total Cycles: 5497

Name:
Sterilizer ID: V2178198
04/11/2025 08:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 71.5 F 0.3
4:00 82.4 F 1.8
6:00 115.2 F 4.3
8:00 144.5 F 7.2
10:00 173.3 F 10.7
12:00 203.0 F 3.4
14:00 207.5 F 4.4
16:00 219.7 F 7.5
18:00 229.4 F 10.9
20:00 230.4 F 6.9
22:00 239.2 F 10.5
24:00 235.8 F 8.0
26:00 247.0 F 13.5
28:00 257.6 F 19.3
30:00 267.7 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.2 F 28.4
2:00 272.3 F 28.9
3:00 271.4 F 28.3
4:00 271.8 F 28.4

Min 271.4 F 28.1
Max 272.4 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 190.2 F 0.0
10:00 180.4 F 0.0
15:00 181.4 F 0.0
20:00 181.0 F 0.0
25:00 180.0 F 0.0
30:00 210.1 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:20
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:09:33
CYCLE COMPLETE

04/11/2025 09:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____