
Midmark M11 - v1.0.5
Total Cycles: 1828

Name: RITTER2
Sterilizer ID: V2313228
12/13/2022 07:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.6 F 0.0
2:00 83.7 F 0.0
4:00 83.8 F 0.0
6:00 110.0 F 0.0
8:00 153.4 F 0.0
10:00 186.6 F 0.7
12:00 217.4 F 3.6
14:00 227.5 F 7.3
16:00 236.2 F 11.2
18:00 231.2 F 7.4
20:00 243.9 F 13.3
22:00 242.0 F 11.7
24:00 240.9 F 10.7
26:00 254.7 F 17.7
28:00 266.4 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.7 F 29.5
2:00 271.8 F 29.0
3:00 272.4 F 28.8
4:00 272.8 F 29.4

Min 270.9 F 28.2
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.9 F 0.0
5:00 206.8 F 0.0
10:00 209.3 F 0.0
15:00 207.9 F 0.0
20:00 202.4 F 0.0
25:00 200.0 F 0.0
30:00 195.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 29:24
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 01:07:18
CYCLE COMPLETE

12/13/2022 08:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____