
Midmark M11 - v1.0.0
Total Cycles: 4186

Name: RITTER 1
Sterilizer ID: V2178198
01/10/2024 09:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.2 F 0.0
2:00 189.5 F 2.2
4:00 198.7 F 6.1
6:00 208.1 F 11.6
8:00 221.3 F 6.0
10:00 232.0 F 10.9
12:00 231.4 F 7.7
14:00 243.8 F 12.5
16:00 243.8 F 12.6
18:00 256.3 F 19.1
20:00 267.3 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.5 F 29.3
2:00 271.5 F 28.6
3:00 272.6 F 29.5
4:00 271.8 F 28.8

Min 271.3 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 218.1 F 0.0
10:00 224.1 F 0.0
15:00 231.0 F 0.0
20:00 223.7 F 0.0
25:00 228.6 F 0.0
30:00 231.4 F 0.0
Drying Complete

FILLING: 2:34
HEATING: 21:28
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:01:35
CYCLE COMPLETE

01/10/2024 10:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____