
Midmark M11 - v1.0.0
Total Cycles: 4529

Name: RITTER 1
Sterilizer ID: V2178198
05/07/2024 08:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 70.0 F 0.0
4:00 78.8 F 0.4
6:00 113.8 F 2.2
8:00 140.0 F 4.5
10:00 163.2 F 7.0
12:00 191.4 F 9.9
14:00 198.4 F 11.2
16:00 207.4 F 4.5
18:00 218.0 F 7.8
20:00 227.8 F 11.3
22:00 229.5 F 6.5
24:00 239.3 F 11.4
26:00 235.3 F 8.6
28:00 248.2 F 14.4
30:00 258.5 F 20.1
32:00 268.2 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.5 F 29.0
2:00 271.4 F 28.3
3:00 272.5 F 29.1
4:00 271.9 F 28.5

Min 271.3 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 187.0 F 0.0
10:00 183.1 F 0.0
15:00 181.5 F 0.0
20:00 181.2 F 0.0
25:00 192.8 F 0.0
30:00 219.5 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 33:23
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:11:43
CYCLE COMPLETE

05/07/2024 09:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____