
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000933
14:41 16/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:51
PRESSURIZING 3:54
STERILIZING 5:24
136.4 C 233kPa 5:24
136.3 C 232kPa 5:54
136.3 C 232kPa 6:24
136.3 C 231kPa 6:54
136.4 C 231kPa 7:24
136.2 C 231kPa 7:54
136.2 C 232kPa 8:24
136.2 C 229kPa 8:54
136.2 C 229kPa 8:54
VENTING 8:55
AIR DRYING 9:48
CYCLE COMPLETE 24:48

Digital Signature #
F4C36BF888F0444B
