
Midmark M11 - v1.0.0
Total Cycles: 2155

Name: RITTER 1
Sterilizer ID: V2178198
07/20/2022 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.3 F 0.0
2:00 183.5 F 1.6
4:00 204.2 F 5.3
6:00 215.5 F 10.0
8:00 224.9 F 5.6
10:00 229.7 F 8.9
12:00 241.2 F 12.0
14:00 238.9 F 10.9
16:00 236.8 F 8.8
18:00 250.4 F 15.6
20:00 261.5 F 22.0
22:00 271.9 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 29.3
2:00 271.8 F 28.5
3:00 272.1 F 29.3
4:00 272.3 F 29.0

Min 271.4 F 28.5
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 223.2 F 0.0
10:00 240.4 F 0.0
15:00 232.7 F 0.0
20:00 227.8 F 0.0
25:00 220.2 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 22:36
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:35
CYCLE COMPLETE

07/20/2022 11:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____