
Midmark M11 - v1.0.0
Total Cycles: 1759

Name: RITTER 1
Sterilizer ID: V2178198
03/29/2022 09:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.5 F 0.0
2:00 152.3 F 0.1
4:00 155.1 F 1.9
6:00 177.7 F 4.9
8:00 201.4 F 9.1
10:00 222.2 F 8.7
12:00 227.2 F 9.0
14:00 238.1 F 10.4
16:00 240.7 F 11.6
18:00 238.1 F 9.6
20:00 252.1 F 16.3
22:00 263.7 F 23.1
24:00 271.4 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.5 F 29.1
2:00 271.9 F 28.4
3:00 271.6 F 28.6
4:00 272.6 F 29.0

Min 271.2 F 28.1
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 214.6 F 0.0
10:00 202.2 F 0.0
15:00 202.8 F 0.0
20:00 230.2 F 0.0
25:00 231.7 F 0.0
30:00 231.3 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 24:08
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:01:58
CYCLE COMPLETE

03/29/2022 10:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____