
Midmark M11 - v1.0.5
Total Cycles: 592

Name: RITTER2
Sterilizer ID: V2313228
12/01/2021 07:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.7 F 0.0
2:00 74.5 F 0.0
4:00 80.2 F 0.1
6:00 110.4 F 0.6
8:00 147.0 F 3.4
10:00 169.5 F 7.3
12:00 186.8 F 11.4
14:00 209.3 F 4.0
16:00 219.6 F 7.7
18:00 230.6 F 11.9
20:00 231.8 F 8.4
22:00 243.2 F 12.4
24:00 243.5 F 12.5
26:00 255.2 F 18.6
28:00 265.5 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 271.7 F 28.9
2:00 271.9 F 29.0
3:00 272.2 F 29.2
4:00 272.6 F 29.5

Min 271.6 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.9 F 0.0
5:00 234.1 F 0.0
10:00 153.4 F 0.0
15:00 166.7 F 0.0
20:00 152.9 F 0.0
25:00 154.9 F 0.0
30:00 152.7 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 29:45
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:07:39
CYCLE COMPLETE

12/01/2021 09:39 AM
ACCEPTED: X
REJECTED:
Operator: U02