

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002468  
16:23 28/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 2:34  
STERILIZING 3:36  
136.0 C 220kPa 3:36  
136.3 C 225kPa 4:07  
136.3 C 222kPa 4:37  
136.3 C 225kPa 5:07  
136.2 C 223kPa 5:37  
136.0 C 222kPa 6:07  
136.4 C 226kPa 6:37  
136.4 C 225kPa 7:07  
136.4 C 225kPa 7:07  
VENTING 7:07  
AIR DRYING 8:03  
CYCLE COMPLETE 8:25  
-----

Digital Signature #  
1E996F5A4F1CA995  
-----