
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002166
10:34 25/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:04
STERILIZING 2:59
136.4 C 223kPa 2:59
136.2 C 224kPa 3:29
136.2 C 224kPa 3:59
136.4 C 225kPa 4:29
136.2 C 225kPa 4:59
136.4 C 226kPa 5:29
136.4 C 226kPa 5:59
136.2 C 222kPa 6:29
136.2 C 222kPa 6:29
VENTING 6:29
AIR DRYING 7:19
CYCLE COMPLETE 37:19

Digital Signature #
2369CFE8D6461B83
