
Midmark M11 - v1.0.0
Total Cycles: 4901

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2024 02:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.2 F 0.0
2:00 155.2 F 0.2
4:00 180.1 F 0.9
6:00 201.2 F 4.2
8:00 215.1 F 7.9
10:00 228.9 F 11.9
12:00 228.8 F 6.7
14:00 239.9 F 11.6
16:00 236.7 F 8.9
18:00 240.8 F 10.1
20:00 248.2 F 14.0
22:00 259.8 F 20.4
24:00 270.0 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.0
1:00 271.6 F 28.2
2:00 272.5 F 28.8
3:00 272.3 F 28.5
4:00 272.0 F 28.2

Min 271.3 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.4 F 0.0
5:00 230.9 F 0.0
10:00 202.5 F 0.0
15:00 194.2 F 0.0
20:00 187.6 F 0.0
25:00 180.9 F 0.0
30:00 179.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:56
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:03:35
CYCLE COMPLETE

09/12/2024 03:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____