
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001702
10:36 08/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:33
STERILIZING 4:39
136.5 C 225kPa 4:39
136.3 C 220kPa 5:09
136.4 C 224kPa 5:39
136.3 C 224kPa 6:09
136.4 C 224kPa 6:39
136.4 C 224kPa 7:09
136.2 C 219kPa 7:39
136.2 C 221kPa 8:09
136.2 C 221kPa 8:09
VENTING 8:09
AIR DRYING 9:04
CYCLE COMPLETE 9:35

Digital Signature #
D62108A86B8B9C38
