
Midmark M11 - v1.0.0
Total Cycles: 1619

Name: RITTER 1
Sterilizer ID: V2178198
02/10/2022 03:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.3 F 0.0
2:00 187.4 F 3.8
4:00 204.2 F 8.5
6:00 241.4 F 12.8
8:00 236.2 F 9.2
10:00 239.2 F 9.7
12:00 244.9 F 12.7
14:00 242.5 F 11.5
16:00 255.6 F 18.2
18:00 267.0 F 25.1
20:00 272.2 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 271.5 F 28.8
2:00 271.4 F 28.8
3:00 271.4 F 28.7
4:00 271.8 F 28.7

Min 271.3 F 28.6
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 244.6 F 0.0
10:00 230.8 F 0.0
15:00 227.4 F 0.0
20:00 223.4 F 0.0
25:00 227.1 F 0.0
30:00 232.3 F 0.0
Drying Complete

FILLING: 2:47
HEATING: 20:07
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

02/10/2022 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____