
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001564
10:42 11/05/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 5:26
STERILIZING 6:54
136.0 C 220kPa 6:54
136.4 C 223kPa 7:24
136.4 C 223kPa 7:54
136.3 C 222kPa 8:25
135.8 C 221kPa 8:55
136.2 C 221kPa 9:25
136.3 C 223kPa 9:55
136.2 C 223kPa 10:24
VENTING 10:25
AIR DRYING 11:27
CYCLE COMPLETE 19:11

Digital Signature #
301FE5B7409D41DE
