
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 001110
14:58 03/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:31
STERILIZING 3:24
136.0 C 220kPa 3:24
136.3 C 225kPa 3:54
136.4 C 226kPa 4:24
136.2 C 222kPa 4:54
136.1 C 223kPa 5:24
136.2 C 224kPa 5:54
136.4 C 222kPa 6:24
136.1 C 221kPa 6:54
136.1 C 221kPa 6:54
VENTING 6:54
AIR DRYING 7:45
CYCLE COMPLETE 20:11

Digital Signature #
9EAA3880B9977EB6
