
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002618
8:53 27/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:32
PRESSURIZING 1:39
STERILIZING 2:48
136.4 C 220kPa 2:48
136.0 C 221kPa 3:18
136.4 C 225kPa 3:48
136.1 C 221kPa 4:18
136.3 C 224kPa 4:48
136.3 C 223kPa 5:18
136.4 C 225kPa 5:48
136.1 C 222kPa 6:18
136.1 C 222kPa 6:18
VENTING 6:18
AIR DRYING 7:09
CYCLE COMPLETE 19:19

Digital Signature #
97F1AEE00DDF70C6
