

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5219

Name: RITTER 1  
Sterilizer ID: V2178198  
01/03/2025 09:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.0 F 0.0  
2:00 77.4 F 0.0  
4:00 91.0 F 0.0  
6:00 124.4 F 0.0  
8:00 151.3 F 0.1  
10:00 179.9 F 0.4  
12:00 199.0 F 2.8  
14:00 209.9 F 5.7  
16:00 219.7 F 8.8  
18:00 229.1 F 12.5  
20:00 229.5 F 7.3  
22:00 240.7 F 12.1  
24:00 237.6 F 9.6  
26:00 238.1 F 9.1  
28:00 248.2 F 14.2  
30:00 259.4 F 20.3  
32:00 269.5 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 272.7 F 28.8  
2:00 272.0 F 28.2  
3:00 271.8 F 28.4  
4:00 272.6 F 28.7  
  
Min 271.4 F 27.9  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.0 F 0.0  
5:00 198.5 F 0.0  
10:00 206.2 F 0.0  
15:00 196.0 F 0.0  
20:00 195.5 F 0.0  
25:00 196.6 F 0.0  
30:00 197.6 F 0.0  
Drying Complete

FILLING: 0:42  
HEATING: 33:07  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:11:24  
CYCLE COMPLETE

01/03/2025 10:46 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_