
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000061
9:07 03/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 3:02
STERILIZING 4:20
136.0 C 220kPa 4:20
136.4 C 225kPa 4:50
136.3 C 222kPa 5:20
135.9 C 222kPa 5:50
136.3 C 224kPa 6:20
136.4 C 223kPa 6:50
136.4 C 225kPa 7:20
136.3 C 223kPa 7:50
136.3 C 223kPa 7:50
VENTING 7:50
AIR DRYING 8:42
CYCLE COMPLETE 15:14

Digital Signature #
53B5DC5E06215B22
