
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000598
12:21 07/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:12
STERILIZING 4:22
136.4 C 223kPa 4:22
136.4 C 222kPa 4:52
136.3 C 224kPa 5:22
136.3 C 224kPa 5:52
136.0 C 223kPa 6:22
136.4 C 222kPa 6:52
136.3 C 224kPa 7:22
136.4 C 224kPa 7:52
136.4 C 224kPa 7:52
VENTING 7:52
AIR DRYING 8:45
CYCLE COMPLETE 8:56

Digital Signature #
C40F8D75F127FE3F
