
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002477
8:03 20/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:17
STERILIZING 4:19
136.4 C 224kPa 4:19
136.3 C 223kPa 4:49
136.2 C 223kPa 5:19
135.9 C 219kPa 5:49
136.4 C 222kPa 6:19
136.2 C 222kPa 6:49
136.3 C 223kPa 7:19
136.3 C 223kPa 7:49
136.3 C 223kPa 7:49
VENTING 7:49
AIR DRYING 8:42
CYCLE COMPLETE 8:44

Digital Signature #
CD3891E918233B9D
